



Test Report

Number: PL 1712064

Page 1 of 3

Applicant:

Address of Applicant:

Date of Receiving Samples: Dec 06, 2017

Testing Period: Dec 07, 2017 to Dec 11, 2017

Place of Testing: Shenzhen, China

Samples Description:

The submitted sample and sample information was/were submitted and identified by/on behalf of client;

Country of Origin:	China	Product Name:	Sunglasses
Country of Destination:	USA	Model No.:	P326.981
Manufacturer/Brand:	Not Provided	Frame Color:	Not Provided
Buyer:	Not Provided	Lens Type:	Uniform Lenses
Lot/Batch Size:	3000 pairs	Lens Color:	Solid Smoke
Sampling Size:	200pcs	P.O. No.:	7198-37BK

Requested Standard:

Test Method:	With Reference to FDA 21 CFR 801.410
Test item:	Impact Resistance

Results/Remarks: Please refer to the following page(s).

Conclusions: The submitted samples are tested and passed the impact resistance on the basis of a statistically significant sampling pursuant to FDA 21 CFR 801. 410.

Issued by stamp



Date of Issued: Dec 12, 2017

For and on behalf of:

Shenzhen Precision Eyewear

Testing & Inspection Services Co., Ltd.

Manager: Wenhua Li

Test Report

Number: PL 1712064

Page 2 of 3

Test Results

P.O. No.	Lens Color	Lot/Batch Size (PCS)	Sampling Size (PCS)	Size of Re.* (PCS)	Size of Ac.* (PCS)	Result(s)
7198-37BK	Solid Smoke	6000	200	0	200	Pass

Acceptance Requirements:

On the basis of a statistically significant sampling pursuant to FDA 21 CFR 801.410, the lenses shall not be fracture.

Sampling Plan From: ANSI /ASQC Standard Z1 .4-2003

(Under 21CFR 801.410(c), it is required to test a statistically significant sample of over-the-counter glass Lenses from each production batch. And FDA accepts the Acceptable Quality (AQL) of 6.5, General Inspection Level II.)

Lot/Batch Size**	Sample Size	Ac/Re*	Lot/Batch Size	Sample Size	Ac/Re
2 to 8	2	0/1	281 to 500	50	7/8
9 to 15	3	0/1	501 to 1200	80	10/11
16 to 25	5	1/2	1201 to 3200	125	14/15
26 to 50	8	1/2	3201 to 10,000	200	21/22
51 to 90	13	2/3	10,001 to 35,000	315	21/22
91 to 150	20	3/4	35,001 to 150,000	500	21/22
151 to 280	32	5/6	150,001 to 500,000	800	21/22

Remark: 1. * Ac=Acceptance; Re=Rejection;

2. ** Generally, a pair of glasses contain 2 pcs of lenses. If the lot size from vendor is refer to the number of glasses, we will multiply it by 2 before sampling.

Test Report

Number: PL 1712064

Page 3 of 3

Appendix: Picture of test samples

-----Report End-----

The test report is effective only with both signature and specialized stamp. The result(s) shown in this report refer only to the sample(s) tested. Without written approval of PEL (Precision Eyewear Lab), this report can't be reproduced except in full.