



INSPECTION REPORT

Report No.: 4478420050716

Sample Name: Nano – Fiber Filter

Sample Model:

Applicant: Xindao B.V.





CHINA CERTIFICATION & INSPECTION GROUP SHENZHEN CO., LTD.
SHENZHEN HUATONGWEI INTERNATIONAL INSPECTION CO.,LTD.



Report No.:4478420050716

Date: 06 04, 2020

Inspection Report

APPLICANT	Xindao B.V.		
ADDRESS	P.O. Box 3082, 2280 GB, Rijswijk, The Netherlands		
SAMPLE NAME	Nano – Fiber Filter		
SAMPLE MODEL	_____		
SAMPLE QUANTITY	1 bag (40 pieces)		
TEST REQUESTED	As specified by client, for details refer to next page(s).		
TEST METHOD	Please refer to next page.		
TEST RESULTS	Please refer to next page(s).		
RECEIVED DATE	05 29, 2020	TESTING PERIOD	05 29, 2020~06 04, 2020
SAMPLE PHOTO:			
1-0(Test sample)		1-1	
			



Report No.:4478420050716

Date: 06 04, 2020

Test Requested :

Test items	Bacterial Filtration Efficiency (BFE), Filtration efficiency
Testing basis	YY 0469-2011 Surgical mask

Test Results:

Test item(s)	Test method(s)	Limit	Measured Values		Judge
Bacterial Filtration Efficiency (BFE) (%)	YY 0469-2011 Appendix B (Bacterial species: Staphylococcus aureus ATCC 6538; Mean particle size:3.0 μ m; Flow rate:28.3L/min; Test area:40cm ² ; Positive Control : 1.94×10 ³ CFU Negative Control:<1CFU)	≥ 95	95.6		Pass
			95.2		
			95.8		
Filtration efficiency (%)	YY 0469-2011	≥ 30 (Non oily particles)	Unpretreated	96.0	Pass
				96.5	
				96.2	

Note: Pass =Up to standard

Fail = Not up to standard

“——” =Not regulated

For the entrusted sample test, the technical responsibilities are undertaken for the submitted samples only.

The test results of this report are only for internal quality control reference.

Approved by:

