

Test Report

Report No.: WP-20086889-JC-01En Page: 1/3

Client: Xindao B.V.

Client Address: P.O.Box 3082,2280 GB,Rijswijk,The Netherlands

The following sample(s) and sample information was/were submitted and confirmed by the client:

Sample Name: Reuseable 2-Ply Cotton Face Mask

Specification: P265.89

Lot No./Type: /

Sample Receiving

Date: 2020.8.27

Test Period: 2020.08.27~2020.09.11

Test Item(s): Please see the next page(s)

Test Result(s): Please see the next page(s)

Approved by: 2020.09.14



Test Report

Report No.: WP-20086889-JC-01En Page: 2/3

1. Test Item:

Test Standard:

Test Result(s):

Sample Name	Testing items	Test Result
Reuseable 2-Ply Cotton Face Mask	Filter efficiency (Salt medium)	95.8%
	Inhalation resistance	122.6Pa
	Exhalation resistance	119.3Pa

Note: Particles with a diameter of $3\pm0.5~\mu m$ are selected.

*** End of the Page ***



Test Report

Report No.: WP-20086889-JC-01En Page: 3/3

Test photo



*** End of the Report***

DISCLAIMER:

- 1. The report is invalid without the stamp of special seal for report or without the signature of the compiler, the inspector and the a pprover.
- 2. No unauthorized changes, additions or deletions shall be made to the report.
- 3. Neither fragmented report nor its incomplete copy shall be deemed valid. The complete copy is invalid without the stamp of special seal for report.
- 4. Any queries on the report shall be presented to Microspectrum Technology Co., Ltd. within 15 working days after receipt of the report.
- 5. The results described here in this report are based on the sample(s) tested. The results are presented to the client for internal uses only and do not constitute any social grounds in the People's Republic of China.
- 6. The client takes full responsible for the truthfulness of the testing sample(s) and information related thereto.